

Test Report No. 3595412-CH02 Date: October 10, 2014 Page 1 of 7

Hearing Aid Amplifier

ON Semiconductor 970 Fraser Drive **Burlington, Ontario** Canada

The following sample(s) was/were submitted

and identified by/on behalf of the client as:

Sample Received Date: 10/7/2014

Testing Period: 10/8/2014 - 10/9/2014

Test Requested Please refer to the result summary.

Test Method & Results Please refer to next page(s).

Result Summary

Test Requested	Comment
European Directive 2011/65/EU Annex II (RoHS); recasting 2002/95/EC	PASS

Model/Part No.: R3110 (laminate based)

Signed for and on behalf of SGS North America, Inc.

Prepared By: Muser

Manoj Aluri

Chemist, Chemistry Laboratory

Veronica Marrero

Laboratory Operations Lead, Chemistry Laboratory



Test Report No. 3595412-CH02 Date: October 10, 2014 Page 2 of 7

European Directive 2011/65/EU Annex II (RoHS); recasting 2002/95/EC

1. Lead, Cadmium, Mercury, Chromium and Bromine content by XRF Screening

Method: With reference to IEC 62321-5:2013. Analysis was performed by XRF Spectroscopy.

Test	Sample Description		Result				Comment
item			Pb	Hg	Cr	Br	Comment
1	Hearing Aid Amplifier – Model/Part No.: R3110 (laminate based) – Side with metal contacts	Х	BL	BL	BL	BL	Refer to chemical result
2	Hearing Aid Amplifier – Model/Part No.: R3110 (laminate based) – Top	Х	BL	BL	BL	BL	Refer to chemical result

Note: 1. BL = Below Limit by XRF analysis

2. **OL** = Over Limit by XRF analysis

3. **X** = Inconclusive (means questionable, need further chemical analysis)

4. # = Insufficient sample for screening test

5. NA = Not Applicable

6. † = With reference to RoHS Annex III Exemption 7(c)-I Electrical and electronic components containing lead in a glass or ceramic other than dielectric ceramic in capacitors, e.g. piezoelectronic devices, or in a glass or ceramic matrix compound.

Remark: 1. Results were obtained by XRF for primary screening, and further chemical testing by ICP (for Cd, Pb, Hg), UV-VIS (for Cr(VI)) and GCMS (for PBBs, PBDEs) are recommended to be performed, if the concentration exceeds the below warning value according to IEC 62321-5:2013.

Element	Unit	Non-metal	Metal	Composite Material
Cd	mg/kg	BL \leq (70-3 σ) $<$ X $<$ (130+3 σ) \leq OL	BL \leq (70-3 σ) $<$ X $<$ (130+3 σ) \leq OL	LOD < X < (150+3σ) ≤ OL
Pb	mg/kg	BL \leq (700-3 σ) $<$ X $<$ (1300+3 σ) \leq OL	BL ≤ (700-3σ) < X < (1300+3σ) ≤ OL	BL ≤ (500-3σ) < X < (1500+3σ) ≤ OL
Hg	mg/kg	BL \leq (700-3 σ) $<$ X $<$ (1300+3 σ) \leq OL	BL ≤ (700-3σ) < X < (1300+3σ) ≤ OL	BL ≤ (500-3σ) < X < (1500+3σ) ≤ OL
Cr	mg/kg	BL ≤ (700-3σ) < X	BL ≤ (700-3σ) < X	BL ≤ (500-3σ) < X
Br	mg/kg	BL ≤ (300-3σ) < X		BL ≤ <i>(</i> 2 <i>50</i> -3σ <i>)</i> < <i>X</i>

BL = Below Limit
OL = Over Limit
X = Inconclusive
LOD = Limit of Detection



Test Report No. 3595412-CH02 Date: October 10, 2014

2. The XRF screening test for RoHS elements – The reading may be different to the actual content in the sample due to non-uniformity composition.

Page 3 of 7

3. Maximum permissible limit:

RoHS Restricted Substances	Maximum Concentration Value (mg/kg) (by weight in homogenous materials)			
Cadmium (Cd)	100			
Lead (Pb)	1000			
Mercury (Hg)	1000			
Hexavalent Chromium (Cr VI)	1000			
Polybrominated biphenyls (PBBs)	1000			
Polybrominated diphenylethers (PBDEs)	1000			



Test Report No. 3595412-CH02 Date: October 10, 2014 Page 4 of 7

2. Cadmium content and Polybromobiphenyl (PBB) and Polybromodiphenyl ether (PBDE) by chemical method

Method: With reference to IEC 62321:2013 and 62321:2008. Cadmium was analyzed by Inductively Coupled Argon Plasma Spectrometry and PBB, PBDE were analyzed by Gas Chromatography – Mass Spectrometry (GC-MS).

Test Item		Result (mg/kg) 1	Detection Limit (mg/kg)	Permissible Limit (mg/kg)
Cadmium	(Cd)	ND	2	100
Comment		PASS		

Sample Description:

1 Hearing Aid Amplifier - Model/Part No.: R3110 (laminate based)- Whole*

Note: 1. mg/kg = milligram per kilogram

2. ND = Not Detected

Remark: * = Sample was tested as a whole because physical separation was not possible



Test Report No. 3595412-CH02 Date: October 10, 2014 Page 5 of 7

Test Item	Result (mg/kg)	Detection Limit (mg/kg)	Permissible Limit (mg/kg)
Sum of PBBs	ND		1000
Monobromobiphenyl	ND	5	
Dibromobiphenyl	ND	5	
Tribromobiphenyl	ND	5	
Tetrabromobiphenyl	ND	5	
Pentabromobiphenyl	ND	5	
Hexabromobiphenyl	ND	5	
Heptabromobiphenyl	ND	5	
Octabromobiphenyl	ND	5	
Nonabromobiphenyl	ND	5	
Decabromobiphenyl	ND	5	
Sum of PBDEs	ND		1000
Monobromodiphenyl ether	ND	5	
Dibromodiphenyl ether	ND	5	
Tribromodiphenyl ether	ND	5	
Tetrabromodiphenyl ether	ND	5	
Pentabromodiphenyl ether	ND	5	
Hexabromodiphenyl ether	ND	5	
Heptabromodiphenyl ether	ND	5	
Octabromodiphenyl ether	ND	5	
Nonabromodiphenyl ether	ND	5	
Decabromodiphenyl ether	ND	5	
Comment	PASS		

Sample Description:

1. Hearing Aid Amplifier - Model/Part No.: R3110 (laminate based)- Whole*

Note: 1. mg/kg = milligram per kilogram

2. ND = Not Detected


Remark: * = Sample was tested as a whole because physical separation was not possible

[&]quot;This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions.aspx and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 45 days only."



Test Report No. 3595412-CH02 Date: October 10, 2014 Page 6 of 7

Flowchart for RoHS:



Note: 1. For Cr(VI), spot test is adopted for metal sample.

- The Cd, Pb and Hg contents test on polymeric samples were dissolved totally by preconditioning method according to above flow chart.
- 3. Cr⁶⁺ is performed only when total Cr is detected



Test Report No. 3595412-CH02 Date: October 10, 2014 Page 7 of 7

Sample Photo:



SGS authenticates the photo on the original report only

*** End of Report ***